

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00314	Calib. Date: 01/16/2012
Certification File No.:	00315	Calib. No.: 00022
Linearity File No.:	00316	Cert. Date: 01/16/2012
Solution File No.:	00325	Cert. No.: 00014
Sequential File No.:	00325	Lin. Date: 01/16/2012
		Lin. No.: 00014
		Soln. Date: 05/25/2012
		Soln. No.: 00106
		File Date: 05/25/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	111095	Expires: 09/15/2013
		Bottle No.: 0448

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:58D		
Control 1 EC	0.101%	11:58D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:59D		
Control 2 EC	0.100%	11:59D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:00D		
Control 3 EC	0.100%	12:01D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:01D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN	First Name: GARY	MI: W
Signature: 	Badge No.: 351	Date: 05/25/2012