

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00210	Calib. Date: 01/06/2009
Certification File No.:	00211	Calib. No.: 00009
Linearity File No.:	00212	Cert. Date: 01/06/2009
Solution File No.:	00220	Cert. No.: 00007
Sequential File No.:	00220	Lin. Date: 01/06/2009
		Lin. No.: 00007
		Soln. Date: 05/28/2009
		Soln. No.: 00059
		File Date: 05/28/2009

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUH S3-0079
Control Solution %:	0.100%	-	Expires: 01/09/2010
Solution Control Lot:	08A049		Bottle No.: 1003

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:54D		
Control 1 EC	0.101%	14:55D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:55D		
Control 2 EC	0.099%	14:56D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:57D		
Control 3 EC	0.100%	14:57D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 258

Date: 05/28/2009