

# Calibrating Unit New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C		Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00235	Calib. Date: 01/06/2009
Certification File No.:	00236	Calib. No.: 00010
Linearity File No.:	00237	Cert. Date: 01/06/2009
Solution File No.:	00259	Cert. No.: 00006
Sequential File No.:	00259	Lin. Date: 01/06/2009
		Lin. No.: 00006
		Soln. Date: 05/28/2009
		Soln. No.: 00067
		File Date: 05/28/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	08L061	Expires: 12/01/2010
		Bottle No.: 1065

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:55D		
Control 1 EC	0.100%	14:56D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:56D		
Control 2 EC	0.100%	14:57D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:57D		
Control 3 EC	0.100%	14:58D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	14:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:59D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: ROGERS

First Name: MICHAEL

MI: J

Signature: \_\_\_\_\_



Badge No.: 258

Date: 05/28/2009