

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00577	Calib. Date: 04/17/2013
Certification File No.:	00578	Calib. No.: 00026
Linearity File No.:	00579	Cert. Date: 04/17/2013
Solution File No.:	00594	Cert. No.: 00016
Sequential File No.:	00594	Lin. Date: 04/17/2013
		Lin. No.: 00016
		Soln. Date: 05/29/2013
		Soln. No.: 00135
		File Date: 05/29/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	12F102	Expires: 06/28/2014
		Bottle No.: 0425

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:58D		
Control 1 EC	0.100%	12:58D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:59D		
Control 2 EC	0.099%	13:00D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:00D		
Control 3 EC	0.099%	13:01D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDSF P2 - 0416 Mad

Changed By:

Last Name: GONCALVES

First Name: MICHELLE

MI: L

Signature: 

Badge No.: 6040

Date: 05/29/2013