

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00411	Calib. No.: 00017
Certification File No.:	00412	Cert. No.: 00011
Linearity File No.:	00413	Lin. No.: 00011
Solution File No.:	00416	Soln. No.: 00100
Sequential File No.:	00416	File Date: 06/02/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	10L084	Expires: 12/16/2012
		Bottle No.: 0139

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:48D		
Control 1 EC	0.098%	11:49D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:50D		
Control 2 EC	0.098%	11:50D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	11:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:51D		
Control 3 EC	0.099%	11:51D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:52D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____

Badge No.: 351

Date: 06/02/2011