

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00394	Calib. Date: 05/01/2014
Certification File No.:	00395	Calib. No.: 00028
Linearity File No.:	00396	Cert. Date: 05/01/2014
Solution File No.:	00399	Cert. No.: 00020
Sequential File No.:	00399	Lin. Date: 05/01/2014
		Lin. No.: 00020
		Soln. Date: 06/03/2014
		Soln. No.: 00138
		File Date: 06/03/2014

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUH S3-0079
Control Solution %: 0.100%		Expires: 11/12/2015
Solution Control Lot: 13K125		Bottle No.: 1341

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:19D		
Control 1 EC	0.100%	16:20D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	16:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:20D		
Control 2 EC	0.099%	16:21D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:22D		
Control 3 EC	0.099%	16:22D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	16:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DROST

First Name: KENNETH

MI: J

Signature: _____



Badge No.: 132

Date: 06/03/2014