

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00673 Calib. Date: 04/07/2014 Calib. No.: 00028  
Certification File No.: 00674 Cert. Date: 04/07/2014 Cert. No.: 00018  
Linearity File No.: 00675 Lin. Date: 04/07/2014 Lin. No.: 00018  
Solution File No.: 00684 Soln. Date: 06/03/2014 Soln. No.: 00149  
Sequential File No.: 00684 File Date: 06/03/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061  
Control Solution %: 0.100% Expires: 11/12/2015  
Solution Control Lot: 13K125 Bottle No.: 1346

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:03D		
Control 1 EC	0.100%	16:03D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:04D		
Control 2 EC	0.099%	16:05D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:05D		
Control 3 EC	0.098%	16:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:07D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: DROST

First Name: KENNETH

MI: J

Signature: \_\_\_\_\_



Badge No.: 132

Date: 06/03/2014