

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00244	Calib. No.: 00015
Certification File No.:	00245	Cert. No.: 00009
Linearity File No.:	00246	Lin. No.: 00009
Solution File No.:	00256	Soln. No.: 00074
Sequential File No.:	00256	File Date: 06/04/2010
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	09I069	Expires: 09/18/2011
		Bottle No.: 0386

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:44D		
Control 1 EC	0.101%	15:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:45D		
Control 2 EC	0.100%	15:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:46D		
Control 3 EC	0.100%	15:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J.

Signature: _____



Badge No.: 258

Date: 06/04/2010