

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00307 Calib. Date: 12/21/2009 Calib. No.: 00012
Certification File No.: 00308 Cert. Date: 12/21/2009 Cert. No.: 00008
Linearity File No.: 00309 Lin. Date: 12/21/2009 Lin. No.: 00008
Solution File No.: 00338 Soln. Date: 06/04/2010 Soln. No.: 00082
Sequential File No.: 00338 File Date: 06/04/2010

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061
Control Solution %: 0.100% Expires: 09/18/2011
Solution Control Lot: 09I069 Bottle No.: 0381

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:38D		
Control 1 EC	0.100%	15:39D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:40D		
Control 2 EC	0.100%	15:40D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	15:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:41D		
Control 3 EC	0.100%	15:42D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J.

Signature: _____



Badge No.: 258

Date: 06/04/2010