

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00831	Calib. No.: 00032
Certification File No.:	00832	Cert. No.: 00022
Linearity File No.:	00833	Lin. No.: 00022
Solution File No.:	00852	Soln. No.: 00180
Sequential File No.:	00852	File Date: 06/09/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	15H140	Expires: 08/04/2017
		Bottle No.: 0090

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:01D		
Control 1 EC	0.102%	10:02D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:02D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:02D		
Control 2 EC	0.102%	10:03D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:03D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:04D		
Control 3 EC	0.101%	10:04D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:04D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 06/09/2016