

# Calibrating Unit

## New Standard Solution Report

|                         |                        |                          |
|-------------------------|------------------------|--------------------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.: ARNK-0074    |
| Location:               | SOUTH BRUNSWICK POLICE |                          |
| Calibration File No.:   | 00743                  | Calib. No.: 00030        |
| Certification File No.: | 00744                  | Cert. No.: 00020         |
| Linearity File No.:     | 00745                  | Lin. No.: 00020          |
| Solution File No.:      | 00767                  | Soln. No.: 00165         |
| Sequential File No.:    | 00767                  | File Date: 06/11/2015    |
| <br>                    |                        |                          |
| Calibrating Unit:       | WET                    | Model No.: CU-34         |
| Control Solution %:     | 0.100%                 | Serial No.: DDUF S3-0060 |
| Solution Control Lot:   | 14H132                 | Expires: 08/22/2016      |
|                         |                        | Bottle No.: 0048         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 17:46D |                |                     |
| Control 1 EC      | 0.101% | 17:47D | 33.9°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.101% | 17:47D | 33.9°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:47D |                |                     |
| Control 2 EC      | 0.100% | 17:48D | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.100% | 17:48D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:49D |                |                     |
| Control 3 EC      | 0.100% | 17:49D | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.100% | 17:49D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:50D |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_

Badge No.: 421

Date: 06/11/2015