

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00424	Calib. No.: 00031
Certification File No.:	00425	Cert. Date: 04/16/2015
Linearity File No.:	00426	Lin. No.: 00022
Solution File No.:	00429	Soln. No.: 00158
Sequential File No.:	00429	File Date: 06/12/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0080
Solution Control Lot:	14H132	Expires: 08/22/2016
		Bottle No.: 0043

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:13D		
Control 1 EC	0.100%	08:13D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:14D		
Control 2 EC	0.100%	08:14D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:15D		
Control 3 EC	0.100%	08:16D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:16D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_

Badge No.: 421

Date: 06/12/2015