

# Calibrating Unit

## New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C		Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00366	Calib. Date: 05/29/2013
Certification File No.:	00367	Cert. No.: 00025
Linearity File No.:	00368	Cert. Date: 05/29/2013
Solution File No.:	00371	Lin. No.: 00017
Sequential File No.:	00371	Soln. No.: 00126
		File Date: 06/27/2013

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUH S3-0080
Control Solution %:	0.100%		Expires: 06/28/2014
Solution Control Lot:	12F102		Bottle No.: 0423

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:53D		
Control 1 EC	0.100%	16:53D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:54D		
Control 2 EC	0.098%	16:54D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:55D		
Control 3 EC	0.098%	16:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:56D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: \_\_\_\_\_

*Sgt. Gary Holsten #351*

Badge No.: 351

Date: 06/27/2013