

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	Calib. No.:	00026
Calibration File No.:	00577	Cert. Date:	04/17/2013
Certification File No.:	00578	Lin. Date:	04/17/2013
Linearity File No.:	00579	Soln. Date:	06/27/2013
Solution File No.:	00607	File Date:	06/27/2013
Sequential File No.:	00607		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0060
Solution Control Lot:	12F102	Expires:	06/28/2014
		Bottle No.:	0422

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:52D		
Control 1 EC	0.099%	16:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:53D		
Control 2 EC	0.099%	16:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:54D		
Control 3 EC	0.100%	16:55D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: Sgt. Gary Holsten #351

Badge No.: 351

Date: 06/27/2013