

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00394	Calib. Date: 05/01/2014
Certification File No.:	00395	Calib. No.: 00028
Linearity File No.:	00396	Cert. Date: 05/01/2014
Solution File No.:	00400	Cert. No.: 00020
Sequential File No.:	00400	Lin. Date: 05/01/2014
		Lin. No.: 00020
		Soln. Date: 07/01/2014
		Soln. No.: 00139
		File Date: 07/01/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	13K125	Expires: 11/12/2015
		Bottle No.: 1340

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:07D		
Control 1 EC	0.100%	21:08D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:08D		
Control 2 EC	0.099%	21:09D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	21:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:10D		
Control 3 EC	0.098%	21:10D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

PH. [Signature] #421

Badge No.: 421

Date: 07/01/2014