

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00287	Calib. Date: 03/22/2011
Certification File No.:	00288	Calib. No.: 00020
Linearity File No.:	00289	Cert. Date: 03/22/2011
Solution File No.:	00301	Cert. No.: 00012
Sequential File No.:	00301	Lin. Date: 03/22/2011
		Lin. No.: 00012
		Soln. Date: 07/05/2011
		Soln. No.: 00091
		File Date: 07/05/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	10L084	Expires: 12/16/2012
		Bottle No.: 0129

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:31D		
Control 1 EC	0.101%	12:32D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	12:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:32D		
Control 2 EC	0.101%	12:33D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	12:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:34D		
Control 3 EC	0.101%	12:34D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	12:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOZSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 07/05/2011