

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00411 Calib. Date: 05/04/2011 Calib. No.: 00017
Certification File No.: 00412 Cert. Date: 05/04/2011 Cert. No.: 00011
Linearity File No.: 00413 Lin. Date: 05/04/2011 Lin. No.: 00011
Solution File No.: 00420 Soln. Date: 07/05/2011 Soln. No.: 00101
Sequential File No.: 00420 File Date: 07/05/2011

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061
Control Solution %: 0.100% Expires: 12/16/2012
Solution Control Lot: 10L084 Bottle No.: 0148

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:24D		
Control 1 EC	0.099%	12:24D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:25D		
Control 2 EC	0.098%	12:25D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:26D		
Control 3 EC	0.099%	12:27D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: 

Badge No.: 351

Date: 07/05/2011