

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00831	Calib. Date: 02/17/2016
Certification File No.:	00832	Cert. No.: 00032
Linearity File No.:	00833	Cert. Date: 02/17/2016
Solution File No.:	00857	Lin. No.: 00022
Sequential File No.:	00857	Lin. Date: 02/17/2016
		Soln. No.: 00181
		Soln. Date: 07/07/2016
		File Date: 07/07/2016

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUF S3-0061
Control Solution %:	0.100%		Expires: 08/04/2017
Solution Control Lot:	15H140		Bottle No.: 0095

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:42D		
Control 1 EC	0.101%	13:43D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:43D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:43D		
Control 2 EC	0.100%	13:44D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:44D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:45D		
Control 3 EC	0.100%	13:45D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:45D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 07/07/2016