

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00743 Calib. Date: 03/19/2015 Calib. No.: 00030
Certification File No.: 00744 Cert. Date: 03/19/2015 Cert. No.: 00020
Linearity File No.: 00745 Lin. Date: 03/19/2015 Lin. No.: 00020
Solution File No.: 00775 Soln. Date: 07/09/2015 Soln. No.: 00166
Sequential File No.: 00775 File Date: 07/09/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060
Control Solution %: 0.100% Expires: 08/22/2016
Solution Control Lot: 14H132 Bottle No.: 0028

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:13D		
Control 1 EC	0.100%	07:14D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:14D		
Control 2 EC	0.100%	07:15D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:16D		
Control 3 EC	0.100%	07:16D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

PH. [Signature] #421

Badge No.: 421

Date: 07/09/2015