

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00743 Calib. Date: 03/19/2015 Calib. No.: 00030  
Certification File No.: 00744 Cert. Date: 03/19/2015 Cert. No.: 00020  
Linearity File No.: 00745 Lin. Date: 03/19/2015 Lin. No.: 00020  
Solution File No.: 00777 Soln. Date: 07/16/2015 Soln. No.: 00167  
Sequential File No.: 00777 File Date: 07/16/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060  
Control Solution %: 0.100% Expires: 08/22/2016  
Solution Control Lot: 14H132 Bottle No.: 0037

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:11D		
Control 1 EC	0.100%	10:12D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:13D		
Control 2 EC	0.100%	10:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:14D		
Control 3 EC	0.100%	10:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_



Badge No.: 421

Date: 07/16/2015