

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00343 Calib. Date: 06/21/2010 Calib. No.: 00013
Certification File No.: 00344 Cert. Date: 06/21/2010 Cert. No.: 00009
Linearity File No.: 00345 Lin. Date: 06/21/2010 Lin. No.: 00009
Solution File No.: 00353 Soln. Date: 07/20/2010 Soln. No.: 00084
Sequential File No.: 00353 File Date: 07/20/2010

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061
Control Solution %: 0.100% Expires: 09/18/2011
Solution Control Lot: 09I069 Bottle No.: 0387

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:15D		
Control 1 EC	0.101%	10:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	10:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:16D		
Control 2 EC	0.101%	10:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	10:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:17D		
Control 3 EC	0.101%	10:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J.

Signature: _____



Badge No.: 258

Date: 07/20/2010