

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00394	Calib. Date: 05/01/2014
Certification File No.:	00395	Calib. No.: 00028
Linearity File No.:	00396	Cert. Date: 05/01/2014
Solution File No.:	00401	Cert. No.: 00020
Sequential File No.:	00401	Lin. Date: 05/01/2014
		Lin. No.: 00020
		Soln. Date: 07/23/2014
		Soln. No.: 00140
		File Date: 07/23/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	13K125	Expires: 11/12/2015
		Bottle No.: 1338

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:58D		
Control 1 EC	0.099%	13:58D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	13:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:59D		
Control 2 EC	0.098%	14:00D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	14:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:00D		
Control 3 EC	0.097%	14:01D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 07/23/2014