

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00673	Calib. Date:	04/07/2014
Certification File No.:	00674	Cert. Date:	04/07/2014
Linearity File No.:	00675	Lin. Date:	04/07/2014
Solution File No.:	00694	Soln. Date:	07/23/2014
Sequential File No.:	00694	File Date:	07/23/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0061
Solution Control Lot:	13K125	Expires:	11/12/2015
		Bottle No.:	1337

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:02D		
Control 1 EC	0.099%	11:02D	34.0°C	AMBIENT AIR CHECK ERR
Control 1 IR	-.---%	11:02D	34.0°C	AMBIENT AIR CHECK ERR
Ambient Air Blank	-.---%	11:03D		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_



Badge No.: 421

Date: 07/23/2014

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00673 Calib. Date: 04/07/2014 Calib. No.: 00028  
Certification File No.: 00674 Cert. Date: 04/07/2014 Cert. No.: 00018  
Linearity File No.: 00675 Lin. Date: 04/07/2014 Lin. No.: 00018  
Solution File No.: 00695 Soln. Date: 07/23/2014 Soln. No.: 00152  
Sequential File No.: 00695 File Date: 07/23/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061  
Control Solution %: 0.100% Expires: 11/12/2015  
Solution Control Lot: 13K125 Bottle No.: 1337

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:27D		
Control 1 EC	0.099%	12:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:28D		
Control 2 EC	0.098%	12:29D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	12:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:30D		
Control 3 EC	0.098%	12:30D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:31D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_



Badge No.: 421

Date: 07/23/2014