

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00577	Calib. Date:	04/17/2013
Certification File No.:	00578	Cert. Date:	04/17/2013
Linearity File No.:	00579	Lin. Date:	04/17/2013
Solution File No.:	00612	Soln. Date:	07/25/2013
Sequential File No.:	00612	File Date:	07/25/2013

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUF S3-0060
Control Solution %:	0.100%			Expires:	06/28/2014
Solution Control Lot:	12F102			Bottle No.:	0421

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:28D		
Control 1 EC	0.099%	06:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	06:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:29D		
Control 2 EC	0.099%	06:30D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:30D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:30D		
Control 3 EC	0.099%	06:31D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	06:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:32D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: Sgt. Gary Holsten #351

Badge No.: 351

Date: 07/25/2013