

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00327	Calib. Date: 07/02/2012
Certification File No.:	00328	Calib. No.: 00023
Linearity File No.:	00329	Cert. Date: 07/02/2012
Solution File No.:	00331	Cert. No.: 00015
Sequential File No.:	00331	Lin. Date: 07/02/2012
		Lin. No.: 00015
		Soln. Date: 07/31/2012
		Soln. No.: 00109
		File Date: 07/31/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	111095	Expires: 09/15/2013
		Bottle No.: 0432

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:47D		
Control 1 EC	0.099%	13:47D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:48D		
Control 2 EC	0.098%	13:48D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:48D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:49D		
Control 3 EC	0.098%	13:49D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:50D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 07/31/2012