

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00498 Calib. Date: 07/02/2012 Calib. No.: 00020  
Certification File No.: 00499 Cert. Date: 07/02/2012 Cert. No.: 00014  
Linearity File No.: 00500 Lin. Date: 07/02/2012 Lin. No.: 00014  
Solution File No.: 00507 Soln. Date: 07/31/2012 Soln. No.: 00118  
Sequential File No.: 00507 File Date: 07/31/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061  
Control Solution %: 0.100% Expires: 09/15/2013  
Solution Control Lot: 111095 Bottle No.: 0427

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:29D		
Control 1 EC	0.099%	12:30D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:30D		
Control 2 EC	0.099%	12:31D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:31D		
Control 3 EC	0.099%	12:32D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: 

Badge No.: 351

Date: 07/31/2012