

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00858 Calib. Date: 07/11/2016 Calib. No.: 00033
Certification File No.: 00859 Cert. Date: 07/11/2016 Cert. No.: 00023
Linearity File No.: 00860 Lin. Date: 07/11/2016 Lin. No.: 00023
Solution File No.: 00863 Soln. Date: 08/02/2016 Soln. No.: 00183
Sequential File No.: 00863 File Date: 08/02/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060
Control Solution %: 0.100% Expires: 08/04/2017
Solution Control Lot: 15H140 Bottle No.: 0097

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:48D		
Control 1 EC	0.103%	12:48D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:49D		
Control 2 EC	0.102%	12:50D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:50D		
Control 3 EC	0.102%	12:51D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:52D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature:  #424

Badge No.: 421

Date: 08/02/2016