

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00424 Calib. Date: 04/16/2015 Calib. No.: 00031
Certification File No.: 00425 Cert. Date: 04/16/2015 Cert. No.: 00022
Linearity File No.: 00426 Lin. Date: 04/16/2015 Lin. No.: 00022
Solution File No.: 00432 Soln. Date: 08/10/2015 Soln. No.: 00160
Sequential File No.: 00432 File Date: 08/10/2015

Calibrating Unit: 0014 Model No.: CU-34 Serial No.: DDUH S3-0080
Control Solution %: 0.100% Expires: 08/22/2016
Solution Control Lot: 14H132 Bottle No.: 0042

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 09:52D | | |
| Control 1 EC | 0.100% | 09:53D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.101% | 09:53D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:53D | | |
| Control 2 EC | 0.100% | 09:54D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.101% | 09:54D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:55D | | |
| Control 3 EC | 0.100% | 09:55D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.101% | 09:55D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:56D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

071 Jesse Blake #421

Badge No.: 421

Date: 08/10/2015