

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00673	Calib. No.: 00028
Certification File No.:	00674	Cert. No.: 00018
Linearity File No.:	00675	Lin. No.: 00018
Solution File No.:	00698	Soln. No.: 00153
Sequential File No.:	00698	File Date: 08/14/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	13K125	Expires: 11/12/2015
		Bottle No.: 1331

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:53D		
Control 1 EC	0.098%	11:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:54D		
Control 2 EC	0.098%	11:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:55D		
Control 3 EC	0.098%	11:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

PH. JB #421

Badge No.: 421

Date: 08/14/2014