

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00257	Calib. Date: 06/21/2010
Certification File No.:	00258	Cert. No.: 00016
Linearity File No.:	00259	Cert. Date: 06/21/2010
Solution File No.:	00265	Lin. Date: 06/21/2010
Sequential File No.:	00265	Lin. No.: 00010
		Soln. Date: 08/19/2010
		Soln. No.: 00077
		File Date: 08/19/2010
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	09I069	Expires: 09/18/2011
		Bottle No.: 0382

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:18D		
Control 1 EC	0.098%	17:19D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	17:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:20D		
Control 2 EC	0.097%	17:20D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	17:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:21D		
Control 3 EC	0.097%	17:21D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	17:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: ROGERS

First Name: MICHAEL

MI: J.

Signature: \_\_\_\_\_

*PFC M. Rogers*

Badge No.: 258

Date: 08/19/2010