

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00343	Calib. Date: 06/21/2010
Certification File No.:	00344	Calib. No.: 00013
Linearity File No.:	00345	Cert. Date: 06/21/2010
Solution File No.:	00358	Cert. No.: 00009
Sequential File No.:	00358	Lin. Date: 06/21/2010
		Lin. No.: 00009
		Soln. Date: 08/19/2010
		Soln. No.: 00085
		File Date: 08/19/2010

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUF S3-0061
Control Solution %:	0.100%		Expires: 09/18/2011
Solution Control Lot:	09I069		Bottle No.: 0377

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:15D		
Control 1 EC	0.101%	17:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	17:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:16D		
Control 2 EC	0.100%	17:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	17:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:17D		
Control 3 EC	0.101%	17:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	17:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: ROGERS

First Name: MICHAEL

MI: J.

Signature: \_\_\_\_\_

*PFC M. Rogers*

Badge No.: 258

Date: 08/19/2010