

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00366	Calib. Date: 05/29/2013
Certification File No.:	00367	Calib. No.: 00025
Linearity File No.:	00368	Cert. Date: 05/29/2013
Solution File No.:	00375	Cert. No.: 00017
Sequential File No.:	00375	Lin. Date: 05/29/2013
		Lin. No.: 00017
		Soln. Date: 08/20/2013
		Soln. No.: 00128
		File Date: 08/20/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0080
Solution Control Lot:	13C111	Expires: 03/12/2015
		Bottle No.: 0135

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:59D		
Control 1 EC	0.100%	14:59D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:00D		
Control 2 EC	0.098%	15:00D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:01D		
Control 3 EC	0.099%	15:02D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____

Sgt. Gary Holsten #351

Badge No.: 351

Date: 08/20/2013