

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00454	Calib. Date: 07/11/2016
Certification File No.:	00455	Calib. No.: 00034
Linearity File No.:	00456	Cert. Date: 07/11/2016
Solution File No.:	00461	Cert. No.: 00025
Sequential File No.:	00461	Lin. Date: 07/11/2016
		Lin. No.: 00025
		Soln. Date: 08/26/2016
		Soln. No.: 00176
		File Date: 08/26/2016

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUH S3-0080
Control Solution %: 0.100%		Expires: 08/04/2017
Solution Control Lot: 15H140		Bottle No.: 0100

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:57D		
Control 1 EC	0.100%	12:57D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:58D		
Control 2 EC	0.099%	12:59D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:59D		
Control 3 EC	0.099%	13:00D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 08/26/2016