

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00858	Calib. Date:	07/11/2016
Certification File No.:	00859	Cert. No.:	00033
Linearity File No.:	00860	Cert. Date:	07/11/2016
Solution File No.:	00865	Lin. No.:	00023
Sequential File No.:	00865	Lin. Date:	07/11/2016
		Soln. No.:	00184
		Soln. Date:	08/26/2016
		File Date:	08/26/2016

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUF S3-0060
Control Solution %:	0.100%			Expires:	08/04/2017
Solution Control Lot:	15H140			Bottle No.:	0099

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:52D		
Control 1 EC	0.099%	12:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:53D		
Control 2 EC	0.099%	12:54D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:54D		
Control 3 EC	0.100%	12:55D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_

Badge No.: 421

Date: 08/26/2016