

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00498 Calib. Date: 07/02/2012 Calib. No.: 00020  
Certification File No.: 00499 Cert. Date: 07/02/2012 Cert. No.: 00014  
Linearity File No.: 00500 Lin. Date: 07/02/2012 Lin. No.: 00014  
Solution File No.: 00519 Soln. Date: 08/28/2012 Soln. No.: 00119  
Sequential File No.: 00519 File Date: 08/28/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061  
Control Solution %: 0.100% Expires: 09/15/2013  
Solution Control Lot: 111095 Bottle No.: 0437

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:50D		
Control 1 EC	0.099%	18:50D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	18:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:51D		
Control 2 EC	0.098%	18:51D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	18:51D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:52D		
Control 3 EC	0.097%	18:53D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.095%	18:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:53D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: 

Badge No.: 351

Date: 08/28/2012