

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0014  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00200 Calib. Date: 08/15/2008 Calib. No.: 00008  
Certification File No.: 00201 Cert. Date: 08/15/2008 Cert. No.: 00006  
Linearity File No.: 00202 Lin. Date: 08/15/2008 Lin. No.: 00006  
Solution File No.: 00204 Soln. Date: 09/13/2008 Soln. No.: 00049  
Sequential File No.: 00204 File Date: 09/13/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060  
Control Solution %: 0.100% Expires: 01/09/2010  
Solution Control Lot: 08A049 Bottle No.: 1011

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:05D		
Control 1 EC	0.101%	23:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	23:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:06D		
Control 2 EC	0.100%	23:07D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	23:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:07D		
Control 3 EC	0.100%	23:08D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J

Signature: \_\_\_\_\_



Badge No.: 258

Date: 09/13/2008

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0014  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00200 Calib. Date: 08/15/2008 Calib. No.: 00008  
Certification File No.: 00201 Cert. Date: 08/15/2008 Cert. No.: 00006  
Linearity File No.: 00202 Lin. Date: 08/15/2008 Lin. No.: 00006  
Solution File No.: 00205 Soln. Date: 10/09/2008 Soln. No.: 00050  
Sequential File No.: 00205 File Date: 10/09/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060  
Control Solution %: 0.100% Expires: 01/09/2010  
Solution Control Lot: 08A049 Bottle No.: 1012

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:52D		
Control 1 EC	0.100%	12:53D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:53D		
Control 2 EC	0.099%	12:54D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:54D		
Control 3 EC	0.100%	12:55D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:56D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: ROGERS

First Name: MICHAEL

MI: J

Signature: \_\_\_\_\_



Badge No.: 258

Date: 10/09/2008

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0014  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00200 Calib. Date: 08/15/2008 Calib. No.: 00008  
Certification File No.: 00201 Cert. Date: 08/15/2008 Cert. No.: 00006  
Linearity File No.: 00202 Lin. Date: 08/15/2008 Lin. No.: 00006  
Solution File No.: 00206 Soln. Date: 11/07/2008 Soln. No.: 00051  
Sequential File No.: 00206 File Date: 11/07/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060  
Control Solution %: 0.100% Expires: 01/09/2010  
Solution Control Lot: 08A049 Bottle No.: 1024

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:02S		
Control 1 EC	0.100%	15:03S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:03S		
Control 2 EC	0.099%	15:04S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:05S		
Control 3 EC	0.099%	15:05S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:06S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: ROGERS

First Name: MICHAEL

MI: J

Signature: \_\_\_\_\_

Badge No.: 258

Date: 11/07/2008

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00200	Calib. Date: 08/15/2008
Certification File No.:	00201	Calib. No.: 00008
Linearity File No.:	00202	Cert. Date: 08/15/2008
Solution File No.:	00208	Cert. No.: 00006
Sequential File No.:	00208	Lin. Date: 08/15/2008
		Lin. No.: 00006
		Soln. Date: 11/24/2008
		Soln. No.: 00052
		File Date: 11/24/2008
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	08A049	Expires: 01/09/2010
		Bottle No.: 1023

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:46S		
Control 1 EC	0.101%	12:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:47S		
Control 2 EC	0.099%	12:48S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:48S		
Control 3 EC	0.100%	12:49S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:50S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: ROGERS

First Name: MICHAEL

MI: J

Signature: \_\_\_\_\_



Badge No.: 258

Date: 11/24/2008

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00200	Calib. Date: 08/15/2008
Certification File No.:	00201	Calib. No.: 00008
Linearity File No.:	00202	Cert. Date: 08/15/2008
Solution File No.:	00209	Cert. No.: 00006
Sequential File No.:	00209	Lin. Date: 08/15/2008
		Lin. No.: 00006
		Soln. Date: 12/24/2008
		Soln. No.: 00053
		File Date: 12/24/2008
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	08A049	Expires: 01/09/2010
		Bottle No.: 1007

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:56S		
Control 1 EC	0.102%	08:57S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:57S		
Control 2 EC	0.100%	08:58S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:59S		
Control 3 EC	0.100%	08:59S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: ROGERS

First Name: MICHEL

MI: J

Signature: \_\_\_\_\_



Badge No.: 258

Date: 12/24/2008