

Calibrating Unit

New Standard Solution Report

| | | | |
|-------------------------|------------------------|--------------|------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARNK-0074 |
| Location: | SOUTH BRUNSWICK POLICE | | |
| Calibration File No.: | 00195 | Calib. Date: | 08/15/2008 |
| Certification File No.: | 00196 | Cert. Date: | 08/15/2008 |
| Linearity File No.: | 00197 | Lin. Date: | 08/15/2008 |
| Solution File No.: | 00204 | Soln. Date: | 09/13/2008 |
| Sequential File No.: | 00204 | File Date: | 09/13/2008 |

| | | | | | |
|-----------------------|--------|------------|-------|-------------|--------------|
| Calibrating Unit: | WET | Model No.: | CU-34 | Serial No.: | DDUH S3-0080 |
| Control Solution %: | 0.100% | | | Expires: | 01/09/2010 |
| Solution Control Lot: | 08A049 | | | Bottle No.: | 1022 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 23:04D | | |
| Control 1 EC | 0.100% | 23:05D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.101% | 23:05D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 23:05D | | |
| Control 2 EC | 0.100% | 23:06D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.101% | 23:06D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 23:06D | | |
| Control 3 EC | 0.100% | 23:07D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.101% | 23:07D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 23:08D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 258

Date: 09/13/2008

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00195 Calib. Date: 08/15/2008 Calib. No.: 00009
Certification File No.: 00196 Cert. Date: 08/15/2008 Cert. No.: 00005
Linearity File No.: 00197 Lin. Date: 08/15/2008 Lin. No.: 00005
Solution File No.: 00210 Soln. Date: 10/09/2008 Soln. No.: 00057
Sequential File No.: 00210 File Date: 10/09/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0080
Control Solution %: 0.100% Expires: 01/09/2010
Solution Control Lot: 08A049 Bottle No.: 1018

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 12:47D | | |
| Control 1 EC | 0.099% | 12:48D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 12:48D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:48D | | |
| Control 2 EC | 0.099% | 12:49D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.101% | 12:49D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:49D | | |
| Control 3 EC | 0.100% | 12:50D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 12:50D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:51D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 258

Date: 10/09/2008

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00195 Calib. Date: 08/15/2008 Calib. No.: 00009
Certification File No.: 00196 Cert. Date: 08/15/2008 Cert. No.: 00005
Linearity File No.: 00197 Lin. Date: 08/15/2008 Lin. No.: 00005
Solution File No.: 00217 Soln. Date: 11/07/2008 Soln. No.: 00058
Sequential File No.: 00217 File Date: 11/07/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0080
Control Solution %: 0.100% Expires: 01/09/2010
Solution Control Lot: 08A049 Bottle No.: 1019

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 14:58S | | |
| Control 1 EC | 0.100% | 14:59S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 14:59S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:59S | | |
| Control 2 EC | 0.100% | 15:00S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.101% | 15:00S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 15:00S | | |
| Control 3 EC | 0.100% | 15:01S | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.101% | 15:01S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 15:02S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 258

Date: 11/07/2008

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00195 Calib. Date: 08/15/2008 Calib. No.: 00009
Certification File No.: 00196 Cert. Date: 08/15/2008 Cert. No.: 00005
Linearity File No.: 00197 Lin. Date: 08/15/2008 Lin. No.: 00005
Solution File No.: 00219 Soln. Date: 11/24/2008 Soln. No.: 00059
Sequential File No.: 00219 File Date: 11/24/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0080
Control Solution %: 0.100% Expires: 01/09/2010
Solution Control Lot: 08A049 Bottle No.: 1013

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 12:51S | | |
| Control 1 EC | 0.101% | 12:51S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 12:51S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:52S | | |
| Control 2 EC | 0.100% | 12:53S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.101% | 12:53S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:53S | | |
| Control 3 EC | 0.101% | 12:54S | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.102% | 12:54S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:55S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ROGERS

First Name: MICHEL

MI: J

Signature: _____



Badge No.: 258

Date: 11/24/2008

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00195 Calib. Date: 08/15/2008 Calib. No.: 00009
Certification File No.: 00196 Cert. Date: 08/15/2008 Cert. No.: 00005
Linearity File No.: 00197 Lin. Date: 08/15/2008 Lin. No.: 00005
Solution File No.: 00230 Soln. Date: 12/24/2008 Soln. No.: 00060
Sequential File No.: 00230 File Date: 12/24/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0080
Control Solution %: 0.100% Expires: 01/09/2010
Solution Control Lot: 08A049 Bottle No.: 1006

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 08:52S | | |
| Control 1 EC | 0.101% | 08:53S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.101% | 08:53S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:53S | | |
| Control 2 EC | 0.101% | 08:54S | 34.0°C | CONTROL TEST FAILED |
| Control 2 IR | 0.111% | 08:54S | 34.0°C | CONTROL TEST FAILED |
| Ambient Air Blank | 0.000% | 08:55S | | |
| Control 3 EC | 0.101% | 08:55S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 08:55S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:56S | | |

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 258

Date: 12/24/2008

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00195 Calib. Date: 08/15/2008 Calib. No.: 00009
Certification File No.: 00196 Cert. Date: 08/15/2008 Cert. No.: 00005
Linearity File No.: 00197 Lin. Date: 08/15/2008 Lin. No.: 00005
Solution File No.: 00231 Soln. Date: 12/24/2008 Soln. No.: 00061
Sequential File No.: 00231 File Date: 12/24/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0080
Control Solution %: 0.100% Expires: 10/03/2009
Solution Control Lot: 07J048 Bottle No.: 1391

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 11:02S | | |
| Control 1 EC | 0.100% | 11:03S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 11:03S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:04S | | |
| Control 2 EC | 0.100% | 11:04S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 11:04S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:05S | | |
| Control 3 EC | 0.100% | 11:06S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.101% | 11:06S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:06S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 258

Date: 12/24/2008