

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00257	Calib. Date: 06/21/2010
Certification File No.:	00258	Calib. No.: 00016
Linearity File No.:	00259	Cert. Date: 06/21/2010
Solution File No.:	00267	Cert. No.: 00010
Sequential File No.:	00267	Lin. Date: 06/21/2010
		Lin. No.: 00010
		Soln. Date: 09/13/2010
		Soln. No.: 00078
		File Date: 09/13/2010
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	09I069	Expires: 09/18/2011
		Bottle No.: 0393

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:47D		
Control 1 EC	0.098%	10:48D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:48D		
Control 2 EC	0.098%	10:49D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:49D		
Control 3 EC	0.098%	10:50D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:51D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 09/13/2010