

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00343 Calib. Date: 06/21/2010 Calib. No.: 00013  
Certification File No.: 00344 Cert. Date: 06/21/2010 Cert. No.: 00009  
Linearity File No.: 00345 Lin. Date: 06/21/2010 Lin. No.: 00009  
Solution File No.: 00363 Soln. Date: 09/13/2010 Soln. No.: 00087  
Sequential File No.: 00363 File Date: 09/13/2010

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061  
Control Solution %: 0.100% Expires: 09/18/2011  
Solution Control Lot: 09I069 Bottle No.: 0388

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:43D		
Control 1 EC	0.101%	10:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:44D		
Control 2 EC	0.101%	10:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:46D		
Control 3 EC	0.101%	10:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: 

Badge No.: 351

Date: 09/13/2010