

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00366	Calib. Date:	05/29/2013
Certification File No.:	00367	Cert. No.:	00025
Linearity File No.:	00368	Cert. Date:	05/29/2013
Solution File No.:	00377	Lin. Date:	05/29/2013
Sequential File No.:	00377	Lin. No.:	00017
		Soln. Date:	09/15/2013
		Soln. No.:	00129
		File Date:	09/15/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0080
Solution Control Lot:	13C111	Expires:	03/12/2015
		Bottle No.:	0149

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:41D		
Control 1 EC	0.101%	18:42D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	18:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:43D		
Control 2 EC	0.100%	18:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	18:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:44D		
Control 3 EC	0.100%	18:44D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	18:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:45D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 09/15/2013