

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00577 Calib. Date: 04/17/2013 Calib. No.: 00026  
Certification File No.: 00578 Cert. Date: 04/17/2013 Cert. No.: 00016  
Linearity File No.: 00579 Lin. Date: 04/17/2013 Lin. No.: 00016  
Solution File No.: 00623 Soln. Date: 09/15/2013 Soln. No.: 00139  
Sequential File No.: 00623 File Date: 09/15/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060  
Control Solution %: 0.100% Expires: 03/12/2015  
Solution Control Lot: 13C111 Bottle No.: 0148

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:35D		
Control 1 EC	0.100%	16:35D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:36D		
Control 2 EC	0.100%	16:37D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:37D		
Control 3 EC	0.100%	16:38D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.101%	16:38D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:38D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: Sgt. Gary Holsten #351

Badge No.: 351

Date: 09/15/2013