

# Calibrating Unit

## New Standard Solution Report

|                         |                        |                          |
|-------------------------|------------------------|--------------------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.: ARTL-0014    |
| Location:               | SOUTH BRUNSWICK POLICE |                          |
| Calibration File No.:   | 00394                  | Calib. No.: 00028        |
| Certification File No.: | 00395                  | Cert. No.: 00020         |
| Linearity File No.:     | 00396                  | Lin. No.: 00020          |
| Solution File No.:      | 00405                  | Soln. No.: 00143         |
| Sequential File No.:    | 00405                  | File Date: 09/15/2014    |
| <br>                    |                        |                          |
| Calibrating Unit:       | WET                    | Model No.: CU-34         |
| Control Solution %:     | 0.100%                 | Serial No.: DDUH S3-0079 |
| Solution Control Lot:   | 13K125                 | Expires: 11/12/2015      |
|                         |                        | Bottle No.: 1333         |

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 08:05D        |                               |                           |
| Control 1 EC      | 0.098%         | 08:05D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.099%         | 08:05D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 08:06D        |                               |                           |
| Control 2 EC      | 0.097%         | 08:06D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.099%         | 08:06D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 08:07D        |                               |                           |
| Control 3 EC      | 0.098%         | 08:08D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.098%         | 08:08D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 08:08D        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_

Badge No.: 421

Date: 09/15/2014