

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00785 Calib. Date: 08/20/2015 Calib. No.: 00031  
Certification File No.: 00786 Cert. Date: 08/20/2015 Cert. No.: 00021  
Linearity File No.: 00787 Lin. Date: 08/20/2015 Lin. No.: 00021  
Solution File No.: 00797 Soln. Date: 09/22/2015 Soln. No.: 00170  
Sequential File No.: 00797 File Date: 09/22/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061  
Control Solution %: 0.100% Expires: 08/22/2016  
Solution Control Lot: 14H132 Bottle No.: 0041

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:21D		
Control 1 EC	0.100%	08:21D	34.1 °C	*** TEST PASSED ***
Control 1 IR	0.100%	08:21D	34.1 °C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:22D		
Control 2 EC	0.100%	08:23D	34.1 °C	*** TEST PASSED ***
Control 2 IR	0.101%	08:23D	34.1 °C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:23D		
Control 3 EC	0.100%	08:24D	34.1 °C	*** TEST PASSED ***
Control 3 IR	0.101%	08:24D	34.1 °C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature:  \_\_\_\_\_

Badge No.: 421

Date: 09/22/2015