

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00454	Calib. Date:	07/11/2016
Certification File No.:	00455	Cert. Date:	07/11/2016
Linearity File No.:	00456	Lin. Date:	07/11/2016
Solution File No.:	00463	Soln. Date:	09/28/2016
Sequential File No.:	00463	File Date:	09/28/2016
Calibrating Unit: WET		Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	4
Solution Control Lot:	16150	Expires:	07/06/2018
		Bottle No.:	1030

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:05D		
Control 1 EC	0.098%	08:06D	34.0°C	AMBIENT AIR CHECK ERR
Control 1 IR	-.---%	08:06D	34.0°C	AMBIENT AIR CHECK ERR
Ambient Air Blank	-.---%	08:06D		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 09/28/2016

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: SOUTH BRUNSWICK POLICE
Serial No.: ARTL-0014
Calibration File No.: 00454 Calib. Date: 07/11/2016 Calib. No.: 00034
Certification File No.: 00455 Cert. Date: 07/11/2016 Cert. No.: 00025
Linearity File No.: 00456 Lin. Date: 07/11/2016 Lin. No.: 00025
Solution File No.: 00464 Soln. Date: 09/28/2016 Soln. No.: 00179
Sequential File No.: 00464 File Date: 09/28/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: 4
Control Solution %: 0.100% Expires: 07/06/2018
Solution Control Lot: 16150 Bottle No.: 1030

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:43D		
Control 1 EC	0.098%	15:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:44D		
Control 2 EC	0.097%	15:44D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:45D		
Control 3 EC	0.098%	15:45D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	15:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 09/28/2016