

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00858 Calib. Date: 07/11/2016 Calib. No.: 00033  
Certification File No.: 00859 Cert. Date: 07/11/2016 Cert. No.: 00023  
Linearity File No.: 00860 Lin. Date: 07/11/2016 Lin. No.: 00023  
Solution File No.: 00872 Soln. Date: 09/23/2016 Soln. No.: 00185  
Sequential File No.: 00872 File Date: 09/23/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060  
Control Solution %: 0.100% Expires: 07/06/2018  
Solution Control Lot: 16150 Bottle No.: 1017

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:59D		
Control 1 EC	0.099%	09:59D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:00D		
Control 2 EC	0.099%	10:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:01D		
Control 3 EC	0.099%	10:02D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_

*Jesse Blake #421*

Badge No.: 421

Date: 09/23/2016