

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00327	Calib. Date: 07/02/2012
Certification File No.:	00328	Calib. No.: 00023
Linearity File No.:	00329	Cert. Date: 07/02/2012
Solution File No.:	00339	Cert. No.: 00015
Sequential File No.:	00339	Lin. Date: 07/02/2012
		Lin. No.: 00015
		Soln. Date: 09/27/2012
		Soln. No.: 00112
		File Date: 09/27/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	111095	Expires: 09/15/2013
		Bottle No.: 0441

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:20D		
Control 1 EC	0.100%	11:21D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:22D		
Control 2 EC	0.098%	11:22D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:23D		
Control 3 EC	0.099%	11:23D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:24D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature:  #357

Badge No.: 351

Date: 09/27/2012