

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00257	Calib. Date:	06/21/2010
Certification File No.:	00258	Cert. Date:	06/21/2010
Linearity File No.:	00259	Lin. Date:	06/21/2010
Solution File No.:	00270	Soln. Date:	10/07/2010
Sequential File No.:	00270	File Date:	10/07/2010

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUH S3-0079
Control Solution %:	0.100%			Expires:	09/18/2011
Solution Control Lot:	09I069			Bottle No.:	0378

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:25D		
Control 1 EC	0.098%	08:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:26D		
Control 2 EC	0.098%	08:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:27D		
Control 3 EC	0.098%	08:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 10/07/2010