

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00343 Calib. Date: 06/21/2010 Calib. No.: 00013  
Certification File No.: 00344 Cert. Date: 06/21/2010 Cert. No.: 00009  
Linearity File No.: 00345 Lin. Date: 06/21/2010 Lin. No.: 00009  
Solution File No.: 00366 Soln. Date: 10/07/2010 Soln. No.: 00088  
Sequential File No.: 00366 File Date: 10/07/2010

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061  
Control Solution %: 0.100% Expires: 09/18/2011  
Solution Control Lot: 09I069 Bottle No.: 0383

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:25D		
Control 1 EC	0.101%	08:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	08:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:27D		
Control 2 EC	0.100%	08:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:28D		
Control 3 EC	0.100%	08:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: \_\_\_\_\_



Badge No.: 351

Date: 10/07/2010