

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00366 Calib. Date: 05/29/2013 Calib. No.: 00025
Certification File No.: 00367 Cert. Date: 05/29/2013 Cert. No.: 00017
Linearity File No.: 00368 Lin. Date: 05/29/2013 Lin. No.: 00017
Solution File No.: 00378 Soln. Date: 10/13/2013 Soln. No.: 00130
Sequential File No.: 00378 File Date: 10/13/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0080
Control Solution %: 0.100% Expires: 03/12/2015
Solution Control Lot: 13C111 Bottle No.: 0143

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:20D		
Control 1 EC	0.101%	20:21D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:22D		
Control 2 EC	0.099%	20:22D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	20:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:23D		
Control 3 EC	0.099%	20:23D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:24D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____

Sgt. Gary Holsten #351

Badge No.: 351

Date: 10/13/2013