

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00577	Calib. Date: 04/17/2013
Certification File No.:	00578	Calib. No.: 00026
Linearity File No.:	00579	Cert. Date: 04/17/2013
Solution File No.:	00628	Cert. No.: 00016
Sequential File No.:	00628	Lin. Date: 04/17/2013
		Lin. No.: 00016
		Soln. Date: 10/13/2013
		Soln. No.: 00140
		File Date: 10/13/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	13C111	Expires: 03/12/2015
		Bottle No.: 0140

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:16D		
Control 1 EC	0.101%	20:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	20:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:17D		
Control 2 EC	0.101%	20:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	20:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:19D		
Control 3 EC	0.101%	20:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	20:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____

Sgt. Gary Holsten #351

Badge No.: 351

Date: 10/13/2013